

Beamline Hazard Analysis - Beamline X9A

LS-BHA-X9A

Revision: 02

Effective: 09/30/11

Page 1 of 1

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

Prepared By: L. Yang

Approved By: Lori Stiegler

Tasks / Equipment	Performed by:		Controls
	Beamline Staff	Users	
Inert compressed gas	Y	Y	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Wear safety glasses with side shields while connecting lines Leak check gas lines
Solvents	Y	N	<ul style="list-style-type: none"> Limit quantities at beamline to < 250 ml Use PPE - safety glasses, nitrile gloves
Ergonomics concerns – moving/lifting	Y	N	<ul style="list-style-type: none"> Be aware of body posture, Ask for help in moving or lifting
Ergonomics concerns – pinch points/sharp edges	Y	Y	<ul style="list-style-type: none"> Warning signs Training
Soldering	Y	N	<ul style="list-style-type: none"> Use a designated area, periodically clean surfaces
High temperature	Y	Y	<ul style="list-style-type: none"> Warning signs Training
Working at heights	Y	N	<ul style="list-style-type: none"> Be aware of body posture, training
Beryllium use	Y	N	<ul style="list-style-type: none"> Handle articles with gloves, Leave area and contact OpCo if damaged
Cutting/razor blades	Y	Y	<ul style="list-style-type: none"> Be aware of hand positioning Use safety knife when possible Consider using cut resistant gloves for frequent cutting tasks